

# Deutsche Akkreditierungsstelle GmbH

# Annex to the Accreditation Certificate D-K-15059-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 07.08.2020

Date of issue: 07.08.2020

Holder of certificate:

## PTW - Freiburg Physikalisch-Technische Werkstätten Dr. Pychlau GmbH Lörracher Straße 7. 79115 Freiburg im Breisgau

Calibration in the fields:

High frequency and radiation quantities

- Ionizing radiation and radioactivity
- Dosimetry
- Radiation protection

#### Abbreviations used: see last page

The certificate together with its annex reflects the status at the time of the date of issue. The current status of the scope of accreditation can be found in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH. https://www.dakks.de/en/content/accredited-bodies-dakks



#### Annex to the accreditation certificate D-K-15059-01-00

#### **Permanent Laboratory**

|   | Calib  | orati                      | ion and Meas  | surement Capabilitie  | s (CMC)   |   |
|---|--|----------------------------|---|---|---|---|
| Measurement<br>quantity /<br>Calibration item |  | Rang                       | je  | Measurement conditions / procedure  | Expanded<br>uncertainty of<br>measurement <sup>1)</sup> | Remarks   |
| <b>Dosimetry</b><br>Air kerma                 | 5 mGy  | to                         | 10 Gy   | X-ray tube voltage,<br>radionuclide resp. radiation<br>quality<br>15 kV to 70 kV                        | 2.0 %   | During gamma<br>radiation indicated<br>dose rates are<br>indications for<br>July 1987, May          |
|   | 2 mGy  | to                         | 10 Gy   | 70 kV to 280 kV<br>20 kV to 50 kV   | 1.9 %<br>2.5 %  | 2008 and/or April<br>2014 according to  |
|   | 100 μGy<br>100 μGy   | to<br>to                   | 10 mGy<br>100 mGy   | (Mammography)<br>40 kV to 150 kV (RAD)  | 2.5 %   | the assigned sources.   |
|   | 100 μGy<br>1 μGy<br>2 μGy                                  | to<br>to<br>to             | 3 Gy<br>5 Gy  | <sup>137</sup> Cs<br><sup>60</sup> Co   | 1.9 %<br>1.2 %  | These are reduced<br>in consequence of  |
| Air kerma rate                                | 50 mGy/min<br>20 mGy/min<br>200 μGy/s                      | to<br>to<br>to             | 500 mGy/min<br>500 mGy/min<br>50 mGy/s                    | 15 kV to 70 kV<br>70 kV to 280 kV<br>20 kV to 50 kV<br>(Mammography)                                    | 2.0 %<br>1.9 %<br>2.5 %                                 | the source strength<br>decrease with the<br>appropriate<br>radioactive half-<br>lives and increased |
|   | 5 μGy/s<br>500 μGy/h<br>1 mGy/h                            | to<br>to<br>to             | 50 mGy/s<br>250 mGy/min<br>500 mGy/min                    | 40 kV to 150 kV (RAD)<br><sup>137</sup> Cs<br><sup>60</sup> Co  | 2.5 %<br>1.9 %<br>1.2 %                                 | if necessary with<br>source change.<br>z <sub>0</sub> : Phantom                                     |
| Ambient<br>equivalent dose                    | 10 μSv<br>3 mSv<br>2 μSv                                   | to<br>to<br>to             | 2 mSv<br>3 Sv<br>5 Sv                                     | 30 kV to 300 kV<br><sup>137</sup> Cs<br><sup>60</sup> Co  | 3.6 %<br>4.6 %<br>4.4 %                                 | surface<br>z <sub>5</sub> : Phantom depth<br>5 cm   |
| Ambient<br>equivalent dose<br>rate            | 1 mSv/h<br>25 mSv/h<br>350 μSv/h<br>0.5 μSv/h<br>500 μSv/h | to<br>to<br>to<br>to<br>to | 400 mSv/h<br>400 mSv/h<br>5 mSv/h<br>10 μSv/h<br>12 mSv/h | 30 kV to 300 kV<br><sup>137</sup> Cs<br><sup>137</sup> Cs<br><sup>137</sup> Cs<br><sup>60</sup> Co      | 3.6 %<br>4.6 %<br>5.3 %<br>7.5 %<br>4.4 %               |   |
| Air kerma length<br>product                   | 700 μGy · cm   | to                         | 700 mGy · cm  | 70 kV to 150 kV   | 2.7%  |   |
| Air kerma length<br>product rate              | 35 μGy · cm/s  |                            | 350 mGy · cm/s  | 70 kV to 150 kV   | 2.7%  |   |
| Absorbed dose to<br>water                     | 10 mGy<br>10 mGy<br>50 mGy                                 | to<br>to<br>to             | 10 Gy<br>10 Gy<br>5 Gy                                    | 10 kV to 100 kV. z <sub>0</sub><br>100 kV to 280 kV. z <sub>5</sub><br><sup>60</sup> Co. z <sub>5</sub> | 3.4 %<br>2.9 %<br>1.1 %                                 |   |
| Absorbed dose<br>rate to water                | 50 mGy/min<br>50 mGy/min<br>50 mGy/min                     | to<br>to<br>to             | 300 mGy/min<br>300 mGy/min<br>300 mGy/min                 | 10 kV to 100 kV. z <sub>0</sub><br>100 kV to 280 kV. z <sub>5</sub><br><sup>60</sup> Co. z <sub>5</sub> | 3.4 %<br>2.9 %<br>1.1 %                                 |   |
|   | > 40 kV  | to                         | 150 kV  |   | 1.2 %   |   |

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of k = 2 unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.



#### Annex to the accreditation certificate D-K-15059-01-00

#### **Permanent Laboratory**

### Calibration and Measurement Capabilities (CMC)

| Measurement<br>quantity /<br>Calibration item |        | Range |        | Measurement conditions / procedure | Expanded<br>uncertainty of<br>measurement <sup>1)</sup> | Remarks  |
|---|--------|-------|--------|------------------------------------|---|--|
| DC voltage                                    | 20 kV  | bis   | 40 kV  | IEC 61676:2002                     | 1.4 %   | For invasive<br>calibration of non-<br>invasive<br>measurement-<br>gadgets |
|   | >40 kV | bis   | 150 kV |                                    | 1.2 %   |  |

#### Abbreviations used:

| CMC | Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten) |
|-----|---|
| IEC | International Electrotechnical Commission                                   |

<sup>&</sup>lt;sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of k = 2 unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.